

CE

IVD

For Professional Use Only

Transport Medium with Mucolytic Agent

Reagent for transportation and storage
of clinical material

Instruction Manual

AmpliSens[®]



Ecoli s.r.o., Studenohorska 12
841 03 Bratislava 47
Slovak Republic
Tel.: +421 2 6478 9336
Fax: +421 2 6478 9040



Federal Budget Institute of
Science "Central Research
Institute for Epidemiology"
3A Novogireevskaya Street
Moscow 111123 Russia

TABLE OF CONTENTS

1. INTENDED USE	3
2. PRINCIPLE.....	3
3. CONTENT	3
4. ADDITIONAL REQUIREMENTS	3
5. GENERAL PRECAUTIONS.....	4
6. SAMPLING AND HANDLING	5
7. WORKING CONDITIONS.....	5
8. PROTOCOL	5
9. TRANSPORTATION.....	5
10. STABILITY AND STORAGE.....	5
11. REFERENCES	6
12. QUALITY CONTROL.....	6
13. KEY TO SYMBOLS USED	7

1. INTENDED USE

Transport Medium with Mucolytic Agent is a reagent intended for transportation and storage of swabs and discharges collected from the urogenital tract, throat, rectum, eye conjunctiva, and erosive-ulcerative lesions of human skin and mucous membranes for subsequent analysis of the material for STIs and other reproductive tract infections by polymerase chain reaction (PCR) and nucleic acid sequence-based amplification (NASBA) with the use of reagent kits manufactured by FBIS CRIE.

2. PRINCIPLE

Transport Medium with Mucolytic Agent is a ready-to-use sterile pink buffer–salt solution supplemented with mucolytic, preservative, and stabilizing agents. The mucolytic agent ensures liquefaction of mucus, provides effective and homogenous mixing of clinical material with the transport medium. The preservative and stabilizing agents prevent the growth of nonspecific microflora and premature lysis of cell, providing long-term stability of RNA/DNA of microorganisms and viruses in a wide temperature range.

3. CONTENT

Transport Medium with Mucolytic Agent is produced in 2 forms:

1 vial of 50 ml, **REF** 952-CE.

100 tubes of 0.5 ml, **REF** 953-CE.

Transport Medium with Mucolytic Agent includes:

<i>Reagent</i>	<i>Description</i>	<i>Volume, ml</i>	<i>Quantity</i>
Transport Medium with Mucolytic Agent	Pink clear liquid	50	1 vial
		or	
		0.5	100 tubes

Transport Medium with Mucolytic Agent is intended for 100 samples.

4. ADDITIONAL REQUIREMENTS

- Disposable powder-free gloves and a laboratory coat.
- Pipettes (adjustable).
- Sterile pipette tips with aerosol filters (up to 1000 µl).
- Disposable polypropylene 2.0-ml tubes.
- Tube racks.
- Disposable sterile probes (tampons or cytobrushes) designed for collecting swabs and discharge from the urogenital tract (cervix, vagina, and urethra), throat, rectum, and erosive-ulcerative lesions of human skin and mucous membranes.

- Reservoir for used tips.

5. GENERAL PRECAUTIONS

The user should always pay attention to the following:

- Use sterile RNase/DNase-free pipette tips with aerosol filters and use a new tip for every procedure.
- Store all extracted positive material (specimens, controls and amplicons) away from all other reagents and add it to the reaction mix in a distantly separated facility.
- Thaw all components thoroughly at room temperature before starting an assay.
- When thawed, mix the components and centrifuge briefly.
- Use disposable protective gloves and laboratory cloths, and protect eyes while samples and reagents handling. Thoroughly wash hands afterwards.
- Do not eat, drink, smoke, apply cosmetics, or handle contact lenses in laboratory work areas.
- Do not use a reagent after its expiration date.
- Dispose of all specimens and unused reagents in accordance with local regulations.
- Samples should be considered potentially infectious and handled in a biological cabinet in compliance with appropriate biosafety practices.
- Clean and disinfect all samples or reagents spills using a disinfectant such as 0.5 % sodium hypochlorite, or other suitable disinfectant.
- Avoid samples and reagents contact with the skin, eyes, and mucous membranes. If these solutions come into contact, rinse the injured area immediately with water and seek medical advice immediately.
- Safety Data Sheets (SDS) are available on request.
- The reagent is intended for single use for analysis of specified number of samples (see the section “Content”).
- The reagent is ready for use in accordance with the Instruction Manual. Use the reagent strictly for intended purpose.
- Use of this product should be limited to personnel trained in DNA amplification techniques.
- Workflow in the laboratory must be one-directional, beginning in the Extraction Area and moving to the Amplification and Detection Area. Do not return samples, equipment and reagents in the area where the previous step was performed.

6. SAMPLING AND HANDLING



Obtaining samples of biological materials for PCR-analysis, transportation, and storage are described in manufacturer's handbook [1]. It is recommended that this handbook is read before starting work.

Transport Medium with Mucolytic Agent is intended for transportation and storage of following clinical material: swabs and discharges collected from the urogenital tract, throat, rectum, eye conjunctiva, and erosive-ulcerative lesions of human skin and mucous membranes.

Storage and transportation of clinical material placed in the **Transport Medium with Mucolytic Agent** (make sure the tube is tightly closed):

- at room temperature (18-25 °C) for up to 28 days;
- at 2-8 °C for up to 3 month;
- at the temperature not more than minus 20 °C for a long time.

7. WORKING CONDITIONS

Transport Medium with Mucolytic Agent should be used at 18–25 °C.

8. PROTOCOL

Omit step 1 if aliquoted form of **Transport Medium with Mucolytic Agent** (100 tubes of 0.5 ml) is used.

1. Dispense 0.5 ml of **Transport Medium with Mucolytic Agent** to 2.0-ml tubes using an aseptic technique. Tightly close the tubes and store them at 2 -25 °C.
2. Prior to opening a tube, make sure that the drops are removed from the tube cap.
3. Place the probe end with clinical material to a tube with **Transport Medium with Mucolytic Agent**, break off the shaft at the scratch mark (if applicable), and recap the tube. If there is no a scratch mark, sink the probe end in the medium, rotate the probe for 5-10 s pressing it to the tube wall, then remove the probe and recap the tube. Mark the tube.

If you have any questions or if you encounter problems, please contact our Authorized representative in the European Community.

9. TRANSPORTATION

Transport Medium with Mucolytic Agent should be transported at 2–25 °C.

10. STABILITY AND STORAGE

Transport Medium with Mucolytic Agent is to be stored at 2–25 °C when not in use.

Transport Medium with Mucolytic Agent is stable until the expiry date stated on the

label. The shelf life of the reagent before and after the first use is the same, unless otherwise stated.














11. REFERENCES

1. Handbook “Sampling, Transportation, and Storage of Clinical Material for PCR Diagnostics” developed by Federal Budget Institute of Science “Central Research Institute for Epidemiology” of Federal Service for Surveillance on Consumers’ Rights Protection and Human Well-Being, Moscow, 2012.

12. QUALITY CONTROL

In compliance with Federal Budget Institute of Science “Central Research Institute for Epidemiology” ISO 13485-Certified Quality Management System each lot of **Transport Medium with Mucolytic Agent** has been tested against predetermined specifications to ensure consistent product quality.

13. KEY TO SYMBOLS USED

	Catalogue number		Caution
	Batch code		Sufficient for
	<i>In vitro</i> diagnostic medical device		Expiration Date
	Version		Consult instructions for use
	Temperature limitation		Upper limit of temperature
	Manufacturer		Date of manufacture
	Authorised representative in the European Community	FBIS CRIE	Federal Budget Institute of Science “Central Research Institute for Epidemiology”

List of Changes Made in the Instruction Manual

VER	Location of changes	Essence of changes
24.06.11 LA	Cover page, text	The name of Institute was changed to Federal Budget Institute of Science “Central Research Institute for Epidemiology”
23.07.12 IV	Content	New complement form was added and described
	Footer	New catalogue number 953-CE was added
27.02.17 PM	Through the text	Corrections according to the template